



### **Competency statements- Brainstem Auditory Evoked Potentials (BAEP)**

These competency statements assume a University education including a bachelor of biomedical science or similar including major components of human anatomy and physiology.

#### **Underpinning Knowledge**

The following areas of knowledge are topics that relate to the competency statements required to perform a BAEP. These areas of knowledge are not included in the statements as topics that require competence in performance but would assist in a better understanding of the competencies required.

- Anatomical structures and function of the central nervous system
- Maturation and development of the central nervous system
- Electrophysiology of the peripheral nervous system
- The neurological examination
- Neuro-imaging techniques
- Diseases of the nervous system including but not limited to
  - acoustic neuroma
  - Multiple Sclerosis
  - sensorineural hearing loss
- Medications used for treatment of diseases of the nervous system
- Verbal and written communication skills
- Health and ethical principles

## **To perform BAEP**

1. Core Knowledge
2. Preparation
3. Patient Care
4. Equipment
5. Electrode application
6. Recording
7. Interpreting the data
8. Completing the test
9. Presenting the data for reporting
10. Managing the recorded data

Appendix

Disclaimer and Copyright

## 1. Core knowledge

The Neurophysiology Health Worker must be able to demonstrate an understanding of specialised knowledge in the following areas:

- Define anatomical structures and function of the
  - auditory nerve pathway
  - brainstem
  - ear
- Describe the effects of diseases of the nervous system on the BAEP including but not limited to
  - acoustic neuroma
  - Multiple Sclerosis
  - sensorineural hearing loss
- Correlate the BAEP waveforms to the appropriate anatomical signal generators
- Identify the use of BAEPs during surgical procedures
- Describe analogue to digital conversion recording techniques
- Identify bandwidth and frequency response characteristics
- Identify and explain the implication and use of frequency filters
- Explain the function and purpose of differential amplifiers
- Define common mode rejection ratio and understand its function and purpose
- Identify the recording parameters and how they differ from display parameters including sensitivity, high frequency filters and low frequency filters
- Identify advantages and disadvantages of different types of electrodes
- Discuss sterilising procedures including high risk infectious diseases according to current Workplace Health and Safety (WH&S) and Infection Control (IC) regulations
- Identify the need to chloride and re-chloride silver electrodes
- Discuss the chemical and electrolytic process to chloride and de-chloride silver electrodes
- Explain the measurement of impedance
- Identify the importance of equal and low impedances in electrode application
- Explain the importance of normative data

## 2. Preparation

The Neurophysiology Health Worker must be competent in the following areas of BAEP preparation:

- Identify the process for patients to attend and leave the clinic
- Prepare consumables appropriately
- Perform routine maintenance of equipment
- Prepare and check equipment is set up according to manufacturer's specifications
- Identify and correct minor equipment faults
- Identify the process for repair of more complex faults
- Prepare the environment according to WH&S regulations
- Identify and remove equipment likely to cause electrical interference
- Register correct data for patient
- Identify sufficient recording space for the recording
- Identify ethical issues that may occur during the recording
- Obtain appropriate patient consent

### 3. Patient Care

The Neurophysiology Health Worker must be competent in the following areas of BAEP patient care:

- Introduce self and others present
- Identify correct patient, correct procedure
- Evaluate clinical and patient information on the request form noting:
  - age of the patient
  - special care requirements
  - obtain relevant clinical history including
    - personal medical history
    - description of symptoms
    - medications
    - family History
  - indication for the test
  - contraindications for the test or part there of
  - identify the stimulation method to best suit the patient (headphones, ear piece)
  - evaluate the appropriate stimulus intensity in relation to the sensory hearing level (sHL)
- Demonstrate appropriate patient interaction
  - according to age, cultural differences and clinical state
  - provide sufficient pre-test information
  - gain sufficient pre-test information
- Explain the procedure including answering questions
- Identify the need to adapt the BAEP procedure according to the information provided
- Position the patient for adequate accessibility, and patient comfort
- Recognise and respond when assistance is required
- Demonstrate patient confidentiality

### 4. Equipment

The Neurophysiology Health Worker must be competent in the following areas of BAEP equipment:

- Electrodes
  - prepare and clean electrodes for use according to WH&S and IC regulations
  - store electrodes appropriately
- Amplifiers
  - identify and explain the implication and use of frequency filters for the BAEP
  - identify the routine acquisition recording parameters for the BAEP
  - identify appropriate sample rates for the BAEP
- Stimulator
  - demonstrate the connection of the stimulator to recording equipment
  - differentiate between different types of stimulation applications
    - rarefaction, condensation, alternating click stimulus
    - presentation methods such as headphones and ear pieces
    - explain the function of masking

- identify the parameters of the click stimulation
  - type
  - polarity
  - rate
  - duration
  - intensity

## 5. Electrode Application

The Neurophysiology Health Worker must be competent in the following areas of BAEP electrode application:

- Accurately apply electrodes according to the 10/20 electrode placement system
  - prepare skin for application of recording electrodes adhering to WH&S and IC regulations
  - explain the importance of good electrode application
  - identify advantages and disadvantages of different types of application – surface electrodes with paste, surface electrodes with Collodion, sub-dermal needle electrodes, other
  - demonstrate appropriate stability of electrode application for the length of recording
  - identify and implement infection control procedures
  - observe and apply standard precautions for contact, droplet and airborne infection risks when applying, removing and cleaning electrodes

## 6. Record the BAEP

The Neurophysiology Health Worker must be competent in the following areas of BAEP recording:

- Connect electrodes to the pre-amplifier
  - connect electrodes to the pre-amplifier according to electrode placement
  - arrange leads and pre-amplifiers to minimise environmental artefacts
- Accurately attach stimulation apparatus to patient
- Identify the appropriate electrode impedance for BAEP
  - define the required impedance level for recording
  - read impedance measurement
  - adjust impedance where appropriate
- Define the machine settings used for BAEP
  - define the machine settings including sensitivity, filters, sweep duration, number of averaged sweeps, automatic rejection
  - explain the relevance of the machine settings to the recording
  - use machine settings according to departmental protocols
  - alter machine settings when appropriate
- Identify the appropriate montage for BAEP
  - identify the montage used for the BAEP
  - discuss advantages and disadvantages of the number of recording channels for the BAEP
- Recognise artefacts
  - identify artefacts and their source

- instrumental and or environmental artefacts
  - physiological artefacts
- eliminate or minimise artefact
- Record the BAEP
  - measure sensory Hearing Threshold (sHL)
  - demonstrate appropriate stimulation techniques for individual sHL
  - demonstrate the appropriate stimulation and masking level for the individual
  - demonstrate appropriate stimulation techniques to determine sHL in patients that cannot actively participate
  - identify appropriate number of sweeps for best signal to noise ratio for BAEP recording
  - demonstrate reproducibility of recorded waveforms
  - follow departmental protocol
  - adjust recording and or stimulation procedure where appropriate
- Annotation of the recording
  - annotate recorded traces to identify appropriate waveform latencies and amplitudes
  - annotate stimulation method used

#### 7. Interpreting the data

The Neurophysiology Health Worker must be competent in the following areas of interpreting the BAEP data:

- Identify the BAEP waveforms
  - identify normal waveforms including latencies and amplitude of the expected BAEP appropriate for age and normative data
  - identify abnormal waveforms including latencies, amplitudes, and side to side differences according to age and normative data
- Correlate the BAEP waveforms to the appropriate anatomical signal generators

#### 8. Completing the BAEP

The Neurophysiology Health Worker must be competent in the following areas of completing the BAEP:

- Check electrode integrity
- Validate recording on completion
- Remove electrodes according to different applications, WH&S and IC regulations and patient comfort
- Remove electrolyte from patient
- Assist patient as required
- Inform patient of the process for obtaining results
- Dispose of materials according to waste management, WH&S and IC regulations
- Clean recording electrodes in accordance with WH&S and IC regulations

9. Presenting BAEP for reporting

The Neurophysiology Health Worker must be competent in the following areas of presenting BAEP for reporting:

- Present recorded traces for interpretation
- Present tabulated data (latencies and amplitudes) for interpretation

10. Managing the recorded data

- Archive recording
- Maintain database of recording

## Appendix – Stakeholders

### Stakeholders

- ANTA Inc. Members
- Document Development Committee
- Document Development Committee Advisory Group
- Other interested parties

### Document Development Committee (2014-2015)

Joanne Wex, Angela Borbelj, Anna Exley, Holly Campbell, Mary Lynch, , Santhi Chigurupati, Malcolm Corkhill, Amy Lofts, Vicky Grant, Samantha Soe.

### Advisory Committee

The document development committee identified a group of key stakeholders to view the draft documents for feedback. The advisory group was made up of technologists, scientists and neurologists working in the neurophysiology industry around Australia. The comments from this group were considered, compared against the reference material and included where appropriate.

### Members Feedback

On completion of the final draft the document was put out to all members of ANTA Inc. for feedback. The comments from members were considered, compared against the reference material and included where appropriate.

### Amendments

2023 July Rebranded to ANSA Inc



**Disclaimer**

- (i) The use of information and data contained within these statements (<http://anta.org.au>) is at your sole risk for ANTA Inc. will not be liable
- (ii) If you rely on the information in these statements you are responsible for ensuring by independent verification their accuracy, currency or completeness. You are also responsible for making sure these statements meet your specific requirements. ANTA Inc. does not warrant or take responsibility for the accuracy, currency or completeness.
- (iii) You acknowledge that such information and materials may contain inaccuracies or errors and we expressly exclude liability for any such inaccuracies or errors to the fullest extent permitted by law.
- (iv) The information and data in these statements is subject to change without notice.
- (v) This document may include links to other websites. These linked websites will have their own terms and conditions of use and you should read and familiarise yourself with these terms and conditions.
- (vi) ANTA Inc. may revise this disclaimer at any time including by updating these statements.

**Copyright ©**

- (i) You are welcome to use these statements and the information contained in this document.
- (ii) All material in this document (<http://anta.org.au>) is protected in Australia under the *Copyright Act 1968 (Cth)*.
- (iii) You are permitted to either electronically save or print out parts of this document for your own information, research or study but are prohibited from using the information for any commercial purpose.
- (iv) You are not permitted to modify copy in any way from the form in which it appears in this document.
- (v) You must not republish any material contained in this document either on another web site, or in any other medium, print, electronic or otherwise without seeking permission to do so. Subject to *Copyright Act 1968* all such use may only be authorised in writing by the copyright owner (ANTA Inc.)
- (vi) Apart from this provision and uses permitted under the *Copyright Act 1968*, all other rights are reserved.